CB101

5 kV Calibration Box



- Gives quick check of insulation tester calibration
- Contains high quality resistors
- Calibration certificate provided

DESCRIPTION

The CB101 is a calibration box suitable for checking the accuracy of insulation testers. Four resistors are connected in the unit to enable a wide range of values to be checked. The high quality resistors, rated to high voltage, can be used to check resistance readings with test volt-ages up to 5 kV d.c. A calibration certificate is provided with each CB101 showing the actual value of each resistance checkpoint.

Specifications

Nominal resistance values: $10 \text{ M}\Omega$, $100 \text{ M}\Omega$,

 $1 \text{ G}\Omega \& 10 \text{ G}\Omega$

Calibration accuracy:1%Nominal value accuracy:2%Voltage coefficient:1ppm/VCalibration temperature:20°C (68°F)

Temperature coefficient: 250 ppm/°C (139 ppm/°F)

Shelf stability Typically: 0.5% per year

Operating temperature: 5°C to 40°C (41°F to 104°F) Storage temperature: -20°C to 60°C (-4°F to 140°F)

Safety: Meets IEC1010-1 (1995)

ORDERING INFORMATION	
Item	Order Code
Calibration Box	CB101

UK

Archcliffe Road Dover CT17 9EN England T +44 (0) 1304 502101 F +44 (0) 1304 207342

UNITED STATES

4271 Bronze Way
Dallas TX 75237-1088 USA
T 800 723 2861 (USA only)
T +1 214 333 3201
F +1 214 331 7399

OTHER TECHNICAL SALES OFFICES

Norristown USA, Toronto CANADA, Mumbai INDIA, Trappes FRANCE, Sydney AUSTRALIA, Madrid SPAIN and the Kingdom of BAHRAIN. Registered to ISO 9001:2000 Reg no. Q 09290 Registered to ISO 14001 Reg no. EMS 61597

CB101_DS_en_V10

www.megger.com Megger is a registered trademark